



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/749,457	12/31/2003	Paul M. Tsou	455770.15022	9830
MARVIN H. KLEINBERG KLEINBERG & LERNER, LLP Suite 1080 2049 Century Park East Los Angeles, CA 90067				
7590 07/25/2008				
EXAMINER				
WERNER, JONATHAN S				
ART UNIT		PAPER NUMBER		
3732				
MAIL DATE		DELIVERY MODE		
07/25/2008		PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/749,457

Applicant(s)

TSOU, PAUL M.

Examiner

JONATHAN WERNER

Art Unit

3732

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 30 November 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-31 is/are pending in the application.
- 4a) Of the above claim(s) 1-18 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 19-31 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/CDC)
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date: _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____
- Paper No(s)/Mail Date: _____

DETAILED ACTION

Drawings

1. The drawings were received on 4/20/07. These drawings are accepted.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 19-20, 23-24, 26, 29-30 are rejected under 35 U.S.C. 102(b) as being anticipated by Kambin (US 5,762,629). Kambin discloses a system for minimal access into soft tissue during spinal surgery, comprising trajectory means which is a first tool (i.e. tapered obturator 20) having a first circumference to seek the appropriate trajectory; tunneling means which is a second tool (i.e. cannula 402) which has an oval cross-section for spreading the worksite it is inserted into (see Figure 23) and a second circumference, said second circumference being larger than that of the first circumference (see Figure 29); and surgical means which is a surgical excision tool (i.e. 350, most notably cutting edge 362 acts as the abrasive tool) having a third circumference which is smaller than the second circumference (see Figure 29).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claims 21-22 and 27-28 rejected under 35 U.S.C. 103(a) as being unpatentable over Kambin in view of Marino (US 6224603 B1). Kambin discloses the surgical access system as described above, but fails to explicitly show that the obturator has a spiral end or that the cannula is beveled. In regard to claims 21 and 27, Marino teaches a similar system which includes a tapered obturator with a spiral end (i.e. 10/20, Figure 2). Therefore, it would have been obvious to one having ordinary skill in the art at the time of Applicant's invention to modify the obturator of Kambin to include a spiral end as taught by Marino in order to facilitate cutting into remaining soft tissue during operation. With respect to claims 22 and 28, Marino teaches the cannula (12) is beveled (at distal end 13). Therefore, it would have been obvious to one having ordinary skill in the art at the time of Applicant's invention to modify the cannula of Kambin to include a beveled end as taught by Marino in order to facilitate cutting into remaining soft tissue during operation.

4. Claims 25 and 31 are rejected under 35 U.S.C. 103(a) as being unpatentable over Kambin in view of Russell et al. (US 5,512,037). Kambin discloses the surgical access system as described above, but fails to explicitly show that the surgical excision

Art Unit: 3732

tool is a flat blade spreader. Russell teaches a surgical access system which includes a cannula (54) through which a flat blade spreader extends (60). Therefore, it would have been obvious to one having ordinary skill in the art at the time of Applicant's invention to modify the excision tool of Kambin to be a flat blade spreader as taught by Russell in order to further retract tissue during a surgical procedure.

Response to Arguments

5. Applicant's arguments with respect to the claims have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Please refer to included form PTO-892 for all additional pertinent prior art related to minimal access surgical systems.

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the

Art Unit: 3732

shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to JONATHAN WERNER whose telephone number is (571)272-2767. The examiner can normally be reached on Monday-Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Cris Rodriguez can be reached on (571) 272-4964. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Jonathan Werner/
Examiner, Art Unit 3732

7/20/08

/Cris L. Rodriguez/
Supervisory Patent Examiner, Art Unit 3732